



MISSOURI DEPARTMENT OF HEALTH  
STATE PUBLIC HEALTH LABORATORY  
DATAMASTER MAINTENANCE REPORT

BREATH ALCOHOL P

JUN -3 20

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN 201288 DATE OF INSPECTION 6-2-2009

LOCATION OF INSTRUMENT (STREET AND CITY) 204 W. Grand Campbell, MO 63933 TIME OF INSPECTION 1314

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED)

☒ COMPUTER

☒ DETECTOR

☒ PROGRAM

☒ FILTERS

☒ HEATERS SAMPLE CHAMBER 49 °C

☒ QUARTZ STANDARD

☒ FLOW DETECTOR

☒ CALIBRATION

☒ PUMP HIGH SPEED

☒ PRINTER

☒ INDICATOR LIGHTS

☒ TIME AND DATE

☒ SIMULATOR TEMPERATURE (34 °C ± 0.2 °C) 34 °C

☒ CALIBRATION CHECK -

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☒ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 ☒ .0977%

TEST 2 ☒ .099%

TEST 3 ☒ .099%

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

☒ NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 1 (.10-.14) 0 (.15-.19) 0 (Over .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

Instrument is within D.O.H. specifications  
Sim Solution Rep to Lot # 08002 exp date Oct 13<sup>th</sup> 2010  
Vapor Alcohol Value .100% +/- 3%

INSPECTING OFFICER

SIGNATURE Weldon Wallace

PRINT NAME Weldon Wallace

TYPE II PERMIT NUMBER/EXPIRATION DATE

820093

03/10/10

TELEPHONE NUMBER

573-246-2511

## CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.  
LOT NUMBER: 08002  
EXPIRATION DATE: October 13, 2010 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 08002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed utilizing a gas chromatograph and found to contain 0.1209 gms/dl wt/vol ethyl alcohol in aqueous solution (i.e. ethanol).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of 0.100% +/-3% when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator.

The date of manufacture for this lot number is October 14, 2008. The expiration date for this lot number is October 13, 2010 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
CAMPBELL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201288  
06/02/09

TESTING OFFICER:  
WALLACE/WELDON/L  
OFFICER I.D.# 1  
PERMIT NUMBER: 820093  
EXPIRATION DATE: 03/10/10  
MISCELLANEOUS DATA:  
MAINTENANCE CHECK

--- SUPERVISOR MODE ---

BLANK TEST	.000	13:23
INTERNAL STANDARD	VERIFIED	13:23
EXTERNAL STANDARD	.097	13:23
BLANK TEST	.000	13:24
EXTERNAL STANDARD	.099	13:24
BLANK TEST	.000	13:25
EXTERNAL STANDARD	.099	13:25
BLANK TEST	.000	13:26

N = 3  
SIM. = .1  
AVG. = .0983

OPERATOR SIGNATURE

CARD STK #  
60036

REORDER ALL SUPPLIES FROM N.P.A.S.  
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (N

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
CAMPBELL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201288  
06/02/09  
13:14

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS  
SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP  
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~"

OPERATOR SIGNATURE

CARD STK #  
60036

REORDER ALL SUPPLIES FROM N.P.A.S.  
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
CAMPBELL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201288  
06/02/09

ARREST TIME: 12:30  
SUBJECT NAME:  
DOE/JON/M  
DOB: 10/10/70 SEX: M  
STATE/D.L.: MO/123456  
ARRESTING OFFICER:  
WALLACE/WELDON/L  
OFFICER I.D.: 1  
TESTING OFFICER:  
WALLACE/WELDON/L  
OFFICER I.D.: 1  
PERMIT NUMBER: 820093  
EXPIRATION DATE: 03/10/10  
MISCELLANEOUS DATA:  
MAINTENANCE CHECK

--- BREATH ANALYSIS ---

BLANK TEST	-000	13:29
INTERNAL STANDARD	VERIFIED	13:29
RADIO INTERFERENCE		

OPERATOR SIGNATURE



CARD STK #  
60036

REORDER ALL SUPPLIES FROM N.P.A.S.  
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)

State of Missouri  
DEPARTMENT OF HEALTH



PERMIT  
TYPE II



WELDON WALLACE

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 03/10/08

Number 820093

Expires 03/10/2010

MO 580-0771 (7-00)

Eric C. Blank  
Director, Department of Health

Director, Department of Health

Lab. 4 (07-8)